

Calibration Certificate of Weighing Balance

ULR No : CC321321000000469F
Page No. : 1 of 2
Certificate No : NC-469
Customer Name : **NSTAR Corp**
34, Pelican Estate
Road No 5
Kathwada GIDC, Kathwada
Ahmedabad - 382430

Certificate Issue Date : October 16, 2021
Calibration Due Date : October 15, 2022

Instrument Detail : **Eletronic Weighing Balance**

Make : SHIMADZU
Model : UW1020H
Serial Number : D482860644
Equipment ID : ML/DWB/007
Avg. Pressure : ==
Avg. Humidity : ==
Calibration Date : October 16, 2021

Capacity : 1020 g
Least Count : 0.001 g
Instrument Location : At Laboratory
Class : I
Stabilization Time (hrs) : 1 hr
Avg. Temperature : ==

References Standards / Masters Used with Traceability					
Sr.No	Denominations	Class	Certified by	Certificate Number	Calibration Validity
1	1 mg to 500 mg	E1	Zwiebel, France	Z20 27234	26-08-2023
2	1 g to 5 kg	E1	Zwiebel, France	Z20 27233	26-08-2023
3					

Important Remarks :

- 01 This certificate is issued for the items submitted to and calibrated by NSTAR Corp Metrology Department.
- 02 Balance location is mentioned in the certificate. Moving the balance after calibration shall be avoided otherwise calibration certificate shall not be accepted as evidence of traceability.
- 03 Thermal stabilization for 1 hr of reference weights have been done before performing calibrations.
- 04 The calibration certificate issued for weighing balances used for scientific or industrial purpose only.
- 05 The weighing balances calibrated for this certificate have been calibrated in accordance with current laboratory procedure NC-SOP-002. The calibration performed meets the criteria as described in the current revisions of NABL 129 and OIML R-76 and ISO/IEC 17025 : 2017.
- 06 Certificate will not be reproduced without written approval of NSTAR Corp. Any corrections, will leads to issue of amended calibration certificate.
- 07 The reported uncertainty is at coverage factor k=2 which corresponds to a coverage probability of approximately 95% for a normal distribution. The contribution of uncertainty originating from the mass standards and balances used and the weighing process are taken into account.



Approved By:
Mr. Naitik Patel

Calibration Certificate of Weighing Balance

ULR No : CC321321000000469F

Page No. : 2 of 2

1. Linearity Test Results						
Sr.No	Denominations	Mass Values of referene weights (g)	Observed Average Mass Value (g)	Error (g)	Expanded Uncertainty at k=2 (g)	Overall Uncertainty (g)
1	20 mg	0.020	0.020	0.000	0.000 676	0.002 539
2	500 mg	0.500	0.500	0.000	0.000 676	
3	1 g	1.000	1.000	0.000	0.000 676	
4	2 g	2.000	2.000	0.000	0.000 676	
5	5 g	5.000	5.000	0.000	0.000 676	
6	10 g	10.000	10.000	0.000	0.000 676	
7	20 g	20.000	20.000	0.000	0.000 676	
8	50 g	50.000	50.000	0.000	0.000 676	
9	100 g	100.000	100.000	0.000	0.000 676	
10	200 g	200.000	200.000	0.000	0.000 677	
11	500 g	500.000	500.000	0.000	0.000 681	
12	1000 g	1000.000	999.999	-0.001	0.000 695	

2. Corner Test / Eccentricity Test Results					
Location	Centre	North	East	South	West
Single Test Weight	500.000	500.000	500.001	500.000	500.001
500 g	500.000	500.001	500.001	500.000	500.001
	500.000				
Average	500.000	500.001	500.001	500.000	500.001
Eccentric Error		-0.000 5	-0.001 0		-0.001 0

3. Repeatability Test Results					
At Half Load	500.000	500.000	500.001	500.000	500.001
500 g	500.000	500.001	500.001	500.000	500.001
At Full Load	1000.000	999.999	1000.000	999.999	999.999
1 kg	1000.000	999.999	999.999	999.999	999.999

4. Hysteresis Test - Not Applicable				
This test is required to be carried out only if the balance is calibrated for the first time or after a major repair.	P1	P2	P3	P4
	Q1	Q2	Q3	Q4
Applicable / Not Applicable	Hysteresis Error			

Document No : NC-FMT-EXE-03 Issue No : 01, Issue Date : 01.06.2020, Revision No : 01, Revision Date : 16.01.2021